

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071224\
 Data File : VV036514.D
 Acq On : 12 Jul 2024 18:07
 Operator : SY/MD
 Sample : P3215-16
 Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4D44

Quant Time: Jul 13 04:11:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 13 04:06:53 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/16/2024
 Supervised By :Semsettin Yesilyurt 07/16/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	480844	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	401976	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	104263	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	113770	16.515	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.040%
7) Chloroethane-d5	1.526	69	104204	17.301	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.200%
11) 1,1-Dichloroethene-d2	2.050	65	58306	16.482	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.920%
21) 2-Butanone-d5	3.863	46	18654m	15.821	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	31.640%
24) Chloroform-d	4.243	84	212232	15.163	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	60.640%
26) 1,2-Dichloroethane-d4	4.937	65	105376	15.180	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	60.720%#
32) Benzene-d6	4.953	84	399379	19.915	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.680%
36) 1,2-Dichloropropane-d6	5.985	67	117856	19.720	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.880%
41) Toluene-d8	7.239	98	316293	17.094	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.360%
43) trans-1,3-Dichloroprop...	7.555	79	32625	13.129	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	52.520%
47) 2-Hexanone-d5	8.037	63	11063	16.666	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	33.340%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	72632	14.430	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	57.720%
66) 1,2-Dichlorobenzene-d4	11.558	152	61038	17.251	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	69.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071224\
Data File : VV036514.D
Acq On : 12 Jul 2024 18:07
Operator : SY/MD
Sample : P3215-16
Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4D44

Quant Time: Jul 13 04:11:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 13 04:06:53 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/16/2024
Supervised By :Semsettin Yesilyurt 07/16/2024

