

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW071624\
 Data File : VW036579.D
 Acq On : 16 Jul 2024 19:54
 Operator : SY/MD
 Sample : P3215-19
 Misc : 5.62g/10.0mL/MSVOA_V/SOIL/B
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4CJ4

Quant Time: Jul 17 02:31:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 17 02:15:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	497474	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	506260	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	186105	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	138476	19.429	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.720%
7) Chloroethane-d5	1.526	69	123226	19.775	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.120%
11) 1,1-Dichloroethene-d2	2.047	65	65201	17.815	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.280%
21) 2-Butanone-d5	3.844	46	25543	20.939	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	41.880%
24) Chloroform-d	4.240	84	282571	19.514	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.040%
26) 1,2-Dichloroethane-d4	4.937	65	143623	19.998	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.000%
32) Benzene-d6	4.953	84	524513	20.767	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.080%
36) 1,2-Dichloropropane-d6	5.982	67	156240	20.757	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.040%
41) Toluene-d8	7.239	98	447293	19.194	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.760%
43) trans-1,3-Dichloroprop...	7.555	79	51428	16.433	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.720%
47) 2-Hexanone-d5	8.030	63	22903	27.396	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.800%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	115502	18.220	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	72.880%
66) 1,2-Dichlorobenzene-d4	11.558	152	152893	24.208	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.840%
Target Compounds						Qvalue
16) Methylene chloride	2.439	84	26506	2.592	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
Data File : VV036579.D
Acq On : 16 Jul 2024 19:54
Operator : SY/MD
Sample : P3215-19
Misc : 5.62g/10.0mL/MSVOA_V/SOIL/B
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4CJ4

Quant Time: Jul 17 02:31:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 02:15:19 2024
Response via : Initial Calibration

