




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
 Data File : VV036571.D
 Acq On : 16 Jul 2024 16:49
 Operator : SY/MD
 Sample : P3238-13
 Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4BF9

Manual IntegrationsAPPROVED

Quant Time: Jul 17 02:29:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 17 02:15:19 2024
 Response via : Initial Calibration

Reviewed By :Romaben Patel 07/17/2024
 Supervised By :Mahesh Dadoda 07/19/2024



TIC: VV036571.D\data.ms

(13) Acetone (T)

2.069min (-0.074) 2.22 ug/L

response 2478

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	14.20	6.62
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
Data File : VV036571.D
Acq On : 16 Jul 2024 16:49
Operator : SY/MD
Sample : P3238-13
Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4BF9

Manual IntegrationsAPPROVED

Quant Time: Jul 17 02:29:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 02:15:19 2024
Response via : Initial Calibration

Reviewed By :Romaben Patel 07/17/2024
Supervised By :Mahesh Dadoda 07/19/2024



TIC: VV036571.D\data.ms

(13) Acetone (T)

2.140min (-0.003) 33.00 ug/L m

response 36785

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	14.20	0.45
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
Data File : VV036571.D
Acq On : 16 Jul 2024 16:49
Operator : SY/MD
Sample : P3238-13
Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4BF9

Manual IntegrationsAPPROVED

Quant Time: Jul 17 02:29:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 02:15:19 2024
Response via : Initial Calibration

Reviewed By :Romaben Patel 07/17/2024
Supervised By :Mahesh Dadoda 07/19/2024



TIC: VV036571.D\data.ms

(21) 2-Butanone-d5 (S)

3.841min (+ 0.007) 15.59 ug/L

response 12292

Ion	Exp%	Act%
46.00	100.00	100.00
77.00	11.40	36.73#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
Data File : VV036571.D
Acq On : 16 Jul 2024 16:49
Operator : SY/MD
Sample : P3238-13
Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4BF9

Manual IntegrationsAPPROVED

Quant Time: Jul 17 02:29:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 02:15:19 2024
Response via : Initial Calibration

Reviewed By :Romaben Patel 07/17/2024
Supervised By :Mahesh Dadoda 07/19/2024



TIC: VV036571.D\data.ms

(21) 2-Butanone-d5 (S)

3.841min (+ 0.007) 18.53 ug/L m

response 14613

Ion	Exp%	Act%
46.00	100.00	100.00
77.00	11.40	30.90#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
Data File : VV036571.D
Acq On : 16 Jul 2024 16:49
Operator : SY/MD
Sample : P3238-13
Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4BF9

Manual IntegrationsAPPROVED

Quant Time: Jul 17 02:29:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 02:15:19 2024
Response via : Initial Calibration

Reviewed By :Romaben Patel 07/17/2024
Supervised By :Mahesh Dadoda 07/19/2024



TIC: VV036571.D\data.ms

(22) 2-Butanone (T)

3.899min (-0.010) 5.02 ug/L

response 4702

Ion	Exp%	Act%
43.00	100.00	100.00
72.00	11.70	7.66
0.00	0.00	0.00
0.00	0.00	0.00


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
Data File : VV036571.D
Acq On : 16 Jul 2024 16:49
Operator : SY/MD
Sample : P3238-13
Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4BF9

Manual IntegrationsAPPROVED

Quant Time: Jul 17 02:29:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 02:15:19 2024
Response via : Initial Calibration

Reviewed By :Romaben Patel 07/17/2024
Supervised By :Mahesh Dadoda 07/19/2024



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
 Data File : VV036571.D
 Acq On : 16 Jul 2024 16:49
 Operator : SY/MD
 Sample : P3238-13
 Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4BF9

Manual IntegrationsAPPROVED

Quant Time: Jul 17 02:29:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 17 02:15:19 2024
 Response via : Initial Calibration

Reviewed By :Romaben Patel 07/17/2024
 Supervised By :Mahesh Dadoda 07/19/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	321557	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	221881	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	43599	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	106547	23.128	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	92.520%	
7) Chloroethane-d5	1.526	69	96862	24.048	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	96.200%	
11) 1,1-Dichloroethene-d2	2.047	65	45415	19.197	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery	=	76.800%	
21) 2-Butanone-d5	3.841	46	14613m	18.533	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	37.060%	
24) Chloroform-d	4.243	84	183269	19.580	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	78.320%	
26) 1,2-Dichloroethane-d4	4.937	65	102966	22.180	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	88.720%	
32) Benzene-d6	4.953	84	325557	29.411	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	117.640%	
36) 1,2-Dichloropropane-d6	5.986	67	103863	31.484	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	125.920%#	
41) Toluene-d8	7.243	98	210128	20.574	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	82.280%	
43) trans-1,3-Dichloroprop...	7.564	79	12707	9.264	ug/L	0.01
Spiked Amount 25.000	Range 30	- 135	Recovery	=	37.040%	
47) 2-Hexanone-d5	8.037	63	8460	23.090	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	46.180%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	66333	23.875	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	95.480%	
66) 1,2-Dichlorobenzene-d4	11.561	152	33493	22.637	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	90.560%	
Target Compounds						
						Qvalue
13) Acetone	2.140	43	36785m	32.998	ug/L	
16) Methylene chloride	2.439	84	49463	7.482	ug/L	96
22) 2-Butanone	3.899	43	18731m	20.014	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071624\
Data File : VV036571.D
Acq On : 16 Jul 2024 16:49
Operator : SY/MD
Sample : P3238-13
Misc : 5.52g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4BF9

Manual IntegrationsAPPROVED

Reviewed By :Romaben Patel 07/17/2024
Supervised By :Mahesh Dadoda 07/19/2024

Quant Time: Jul 17 02:29:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071124SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 02:15:19 2024
Response via : Initial Calibration

