

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072624\
 Data File : VW036703.D
 Acq On : 25 Jul 2024 13:31
 Operator : SY/MD
 Sample : P3276-08
 Misc : 9.02g/10mL/MSVOA_V/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Jul 26 05:01:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072224SMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 25 00:49:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	434968	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	357304	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	97039	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	49675	10.427	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	41.720%
7) Chloroethane-d5	1.529	69	57417	13.344	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.360%
11) 1,1-Dichloroethene-d2	2.050	65	26404	10.482	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	41.920%#
21) 2-Butanone-d5	3.815	46	18394	16.544	ug/L	-0.04
Spiked Amount	50.000	Range	20 - 135	Recovery	=	33.080%
24) Chloroform-d	4.246	84	193215	16.796	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	67.200%
26) 1,2-Dichloroethane-d4	4.941	65	102838	18.120	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	72.480%
32) Benzene-d6	4.957	84	264572	15.623	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	62.480%
36) 1,2-Dichloropropane-d6	5.986	67	100998	20.932	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.720%
41) Toluene-d8	7.243	98	192198	11.893	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	47.560%
43) trans-1,3-Dichloroprop...	7.558	79	32702	15.246	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	61.000%
47) 2-Hexanone-d5	8.034	63	26661	41.519	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.040%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	106852	25.408	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.640%
66) 1,2-Dichlorobenzene-d4	11.558	152	80944	23.895	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.600%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.278	62	34869	4.968	ug/L	99
10) 1,1,2-Trichloro-1,2,2-...	2.063	101	3872	0.511	ug/L #	28
12) 1,1-Dichloroethene	2.063	96	1308906	195.059	ug/L #	81
16) Methylene chloride	2.442	84	23416	2.893	ug/L	94
17) trans-1,2-Dichloroethene	2.693	96	11089	1.636	ug/L	99
20) cis-1,2-Dichloroethene	3.806	96	1890273	267.089	ug/L	99
34) Trichloroethene	5.825	95	1172078	176.507	ug/L	98
35) Methylcyclohexane	5.989	83	27439	3.103	ug/L #	24
61) 1,2,3-Trichloropropane	11.181	75	9036	5.282	ug/L #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072624\
Data File : VV036703.D
Acq On : 25 Jul 2024 13:31
Operator : SY/MD
Sample : P3276-08
Misc : 9.02g/10mL/MSVOA_V/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Jul 26 05:01:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072224SMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 25 00:49:57 2024
Response via : Initial Calibration

