

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072718\
 Data File : VV006720.D
 Acq On : 27 Jul 2018 11:15
 Operator : SY/MD
 Sample : VV0727WBL01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK72

Quant Time: Jul 28 02:34:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 28 02:24:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	272229	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	250414	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	105670	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	84379	5.96	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	119.20%
7) Chloroethane-d5	1.58	69	64383	5.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.80%
11) 1,1-Dichloroethene-d2	2.13	63	113924	4.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.00%
20) 2-Butanone-d5	3.97	46	202832	45.82	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.64%
24) Chloroform-d	4.40	84	141536	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.40%
26) 1,2-Dichloroethane-d4	5.09	65	71801	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.80%
32) Benzene-d6	5.10	84	302778	5.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.60%
36) 1,2-Dichloropropane-d6	6.12	67	96491	5.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.00%
41) Toluene-d8	7.36	98	260544	5.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.80%
43) trans-1,3-Dichloropropene-	7.67	79	30920	5.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	109.00%
46) 2-Hexanone-d5	8.14	63	162192	45.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.30%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	66925	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	91910	5.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072718\
Data File : VV006720.D
Acq On : 27 Jul 2018 11:15
Operator : SY/MD
Sample : VV0727WBL01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK72

Quant Time: Jul 28 02:34:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 28 02:24:46 2018
Response via : Initial Calibration

