

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072718\
 Data File : VV006722.D
 Acq On : 27 Jul 2018 12:50
 Operator : SY/MD
 Sample : J4148-02
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0SW7

Quant Time: Jul 28 02:44:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 28 02:24:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	254667	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	235070	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	97881	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83238	6.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	125.60%
7) Chloroethane-d5	1.58	69	63060	6.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.20%
11) 1,1-Dichloroethene-d2	2.13	63	110348	4.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.80%
20) 2-Butanone-d5	3.95	46	204169	49.30	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.60%
24) Chloroform-d	4.41	84	138751	5.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.60%
26) 1,2-Dichloroethane-d4	5.09	65	70751	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.60%
32) Benzene-d6	5.10	84	292907	6.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.20%
36) 1,2-Dichloropropane-d6	6.12	67	94837	5.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	116.20%
41) Toluene-d8	7.37	98	251621	5.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.00%
43) trans-1,3-Dichloropropene-	7.67	79	29411	5.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	110.40%
46) 2-Hexanone-d5	8.14	63	158797	47.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.18%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	66487	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	89711	6.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	123.00%#

Target Compounds					Ovalue
13) Acetone	2.20	43	9343	3.546 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072718\
Data File : VV006722.D
Acq On : 27 Jul 2018 12:50
Operator : SY/MD
Sample : J4148-02
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0SW7

Quant Time: Jul 28 02:44:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 28 02:24:46 2018
Response via : Initial Calibration

