

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072718\
 Data File : VV006727.D
 Acq On : 27 Jul 2018 15:05
 Operator : SY/MD
 Sample : J4190-06
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BE746

Quant Time: Jul 28 02:58:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 28 02:24:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	150467	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	138016	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	53953	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	50167	6.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	128.20%
7) Chloroethane-d5	1.58	69	39045	6.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	127.00%
11) 1,1-Dichloroethene-d2	2.13	63	67693	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.20%
20) 2-Butanone-d5	3.95	46	133430	54.53	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.06%
24) Chloroform-d	4.41	84	85880	5.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.80%
26) 1,2-Dichloroethane-d4	5.09	65	43956	6.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	122.80%
32) Benzene-d6	5.10	84	175406	6.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	123.60%
36) 1,2-Dichloropropane-d6	6.12	67	57595	6.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	120.20%
41) Toluene-d8	7.37	98	145027	5.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.00%
43) trans-1,3-Dichloropropene-	7.67	79	17366	5.55	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	111.00%
46) 2-Hexanone-d5	8.14	63	91421	46.17	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.34%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	42492	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	49974	6.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.20%#

Target Compounds					Ovalue
13) Acetone	2.19	43	10200	6.552 ug/L	99
17) Methyl tert-butyl Ether	2.82	73	9461	0.391 ug/L	99
25) Chloroform	4.43	83	7319	0.396 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072718\
Data File : VV006727.D
Acq On : 27 Jul 2018 15:05
Operator : SY/MD
Sample : J4190-06
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BE746

Quant Time: Jul 28 02:58:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 28 02:24:46 2018
Response via : Initial Calibration

