

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072718\
 Data File : VV006730.D
 Acq On : 27 Jul 2018 16:26
 Operator : SY/MD
 Sample : J4209-08
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AC3

Quant Time: Jul 28 03:38:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 28 02:24:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	185875	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	172072	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	68258	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57693	5.96	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	119.20%
7) Chloroethane-d5	1.58	69	45270	5.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.20%
11) 1,1-Dichloroethene-d2	2.13	63	79510	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.80%
20) 2-Butanone-d5	3.96	46	151463	50.11	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.22%
24) Chloroform-d	4.41	84	100157	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
26) 1,2-Dichloroethane-d4	5.09	65	51865	5.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.20%
32) Benzene-d6	5.10	84	206709	5.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.80%
36) 1,2-Dichloropropane-d6	6.12	67	68043	5.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.80%
41) Toluene-d8	7.37	98	172148	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.40%
43) trans-1,3-Dichloropropene-	7.67	79	21221	5.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	108.80%
46) 2-Hexanone-d5	8.14	63	110586	44.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.58%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	49505	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	60235	5.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.40%

Target Compounds					Ovalue
13) Acetone	2.20	43	3568	1.855 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072718\
Data File : VV006730.D
Acq On : 27 Jul 2018 16:26
Operator : SY/MD
Sample : J4209-08
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AC3

Quant Time: Jul 28 03:38:27 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 28 02:24:46 2018
Response via : Initial Calibration

