

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072718\
 Data File : VV006735.D
 Acq On : 27 Jul 2018 20:02
 Operator : SY/MD
 Sample : J4207-01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JJ1G1

Quant Time: Jul 28 03:52:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 28 03:46:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	217810	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	203598	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	84875	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	66447	5.86	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.20%
7) Chloroethane-d5	1.57	69	52817	5.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	118.60%
11) 1,1-Dichloroethene-d2	2.13	63	89425	4.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.40%
20) 2-Butanone-d5	3.95	46	182751	51.60	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.20%
24) Chloroform-d	4.40	84	117881	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.80%
26) 1,2-Dichloroethane-d4	5.09	65	60682	5.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.00%
32) Benzene-d6	5.10	84	242350	5.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.80%
36) 1,2-Dichloropropane-d6	6.12	67	79202	5.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.00%
41) Toluene-d8	7.36	98	201585	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.20%
43) trans-1,3-Dichloropropene-	7.67	79	24963	5.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	108.20%
46) 2-Hexanone-d5	8.14	63	141439	48.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.84%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	60328	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	73267	5.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.80%

Target Compounds					Ovalue
13) Acetone	2.19	43	5102	2.26 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072718\
Data File : VV006735.D
Acq On : 27 Jul 2018 20:02
Operator : SY/MD
Sample : J4207-01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JJ1G1

Quant Time: Jul 28 03:52:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 28 03:46:17 2018
Response via : Initial Calibration

