

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072718\
 Data File : VV006738.D
 Acq On : 27 Jul 2018 21:23
 Operator : SY/MD
 Sample : J4207-04
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JJ1G4

Quant Time: Jul 28 03:56:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 28 03:46:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	210533	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	196554	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	87847	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	62423	5.70	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.00%
7) Chloroethane-d5	1.58	69	48418	5.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.60%
11) 1,1-Dichloroethene-d2	2.13	63	82770	3.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.00%
20) 2-Butanone-d5	3.96	46	171970	50.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.46%
24) Chloroform-d	4.41	84	111227	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
26) 1,2-Dichloroethane-d4	5.09	65	57118	5.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.00%
32) Benzene-d6	5.10	84	223621	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.80%
36) 1,2-Dichloropropane-d6	6.12	67	75315	5.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.40%
41) Toluene-d8	7.37	98	183371	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.80%
43) trans-1,3-Dichloropropene-	7.67	79	22724	5.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.00%
46) 2-Hexanone-d5	8.14	63	134423	47.67	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.34%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	56810	5.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	72847	5.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.20%


Target Compounds					Ovalue
13) Acetone	2.21	43	5398	2.48 ug/L	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072718\
Data File : VV006738.D
Acq On : 27 Jul 2018 21:23
Operator : SY/MD
Sample : J4207-04
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JJ1G4

Quant Time: Jul 28 03:56:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 28 03:46:17 2018
Response via : Initial Calibration





#13

Acetone

Concen: 2.48 ug/L

RT: 2.21 min Scan# 409

Delta R.T. 0.00 min

Lab File: VV006738.D

Acq: 27 Jul 2018 21:23

Instrument :
MSVOA_V
ClientSampleId :
JJ1G4

Tot Ion: 43 Resp: 5398

Ion Ratio Lower Upper

43 100

58 35.6 0.0 71.4

