

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080122\
 Data File : VW027147.D
 Acq On : 02 Aug 2022 00:09
 Operator : SY/MD
 Sample : N3956-04DL 20X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXF03DL

Quant Time: Aug 02 00:38:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 02 00:27:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	618968	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	598388	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	253699	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	267883	66.780	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	133.560%
7) Chloroethane-d5	1.558	69	122516	40.814	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.620%
11) 1,1-Dichloroethene-d2	2.098	63	335860	43.968	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.940%
21) 2-Butanone-d5	3.876	46	316422	119.809	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.810%
24) Chloroform-d	4.339	84	445053	54.887	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.780%
26) 1,2-Dichloroethane-d4	5.024	65	269937	58.675	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.360%
32) Benzene-d6	5.044	84	925955	58.478	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	116.960%
36) 1,2-Dichloropropane-d6	6.063	67	290451	59.297	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	118.600%
41) Toluene-d8	7.310	98	786746	55.114	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.220%
43) trans-1,3-Dichloroprop...	7.619	79	123000	57.216	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	114.440%
47) 2-Hexanone-d5	8.088	63	267906	141.672	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	141.670%#
56) 1,1,2,2-Tetrachloroeth...	10.214	84	406287	55.489	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	110.980%
66) 1,2-Dichlorobenzene-d4	11.622	152	281808	63.205	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	126.400%#
Target Compounds					Qvalue	
33) Benzene	5.095	78	256250	13.962	ug/L	100
51) Chlorobenzene	8.883	112	13380	1.021	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027147.D
Acq On : 02 Aug 2022 00:09
Operator : SY/MD
Sample : N3956-04DL 20X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXF03DL

Quant Time: Aug 02 00:38:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 02 00:27:15 2022
Response via : Initial Calibration

