

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080122\
 Data File : VW027150.D
 Acq On : 02 Aug 2022 01:22
 Operator : SY/MD
 Sample : N3956-09DL 40X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXF06DL

Quant Time: Aug 02 04:00:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 02 00:27:15 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.613	114	645468	50.000	ug/L	0.00
28) Chlorobenzene-d5		8.850	117	617411	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.246	152	272739	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.304	65	280461	67.045	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	134.080%	
7) Chloroethane-d5		1.558	69	127151	40.619	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.240%	
11) 1,1-Dichloroethene-d2		2.098	63	357800	44.917	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.840%	
21) 2-Butanone-d5		3.876	46	308407	111.980	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.980%	
24) Chloroform-d		4.339	84	454770	53.783	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.560%	
26) 1,2-Dichloroethane-d4		5.024	65	274618	57.242	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.480%	
32) Benzene-d6		5.043	84	961777	58.869	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.740%	
36) 1,2-Dichloropropane-d6		6.063	67	297447	58.855	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	117.700%	
41) Toluene-d8		7.313	98	826808	56.136	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.280%	
43) trans-1,3-Dichloroprop...		7.619	79	125336	56.506	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	113.020%	
47) 2-Hexanone-d5		8.088	63	253923	130.140	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	130.140%#	
56) 1,1,2,2-Tetrachloroeth...		10.214	84	409247	54.171	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.340%	
66) 1,2-Dichlorobenzene-d4		11.622	152	301834	62.970	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	125.940%#	
Target Compounds							
13) Acetone		2.169	43	27947	12.017	ug/L	95
33) Benzene		5.092	78	746848	39.440	ug/L	100
42) Toluene		7.387	91	114163	5.844	ug/L	99
51) Chlorobenzene		8.879	112	1267363	93.748	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027150.D
Acq On : 02 Aug 2022 01:22
Operator : SY/MD
Sample : N3956-09DL 40X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXF06DL

Quant Time: Aug 02 04:00:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 02 00:27:15 2022
Response via : Initial Calibration

