

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080122\
 Data File : VW027153.D
 Acq On : 02 Aug 2022 02:35
 Operator : SY/MD
 Sample : N3880-16 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUG9

Quant Time: Aug 02 04:10:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 02 00:27:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	611048	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	587053	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	248468	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	271989	68.682	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	137.360%#	
7) Chloroethane-d5	1.555	69	115187	38.870	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	77.740%	
11) 1,1-Dichloroethene-d2	2.095	63	341377	45.270	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	90.540%	
21) 2-Butanone-d5	3.876	46	303598	116.444	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	116.440%	
24) Chloroform-d	4.339	84	446464	55.775	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	111.540%	
26) 1,2-Dichloroethane-d4	5.024	65	265654	58.493	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	116.980%	
32) Benzene-d6	5.043	84	931308	59.952	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	119.900%	
36) 1,2-Dichloropropane-d6	6.063	67	288311	59.997	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	120.000%	
41) Toluene-d8	7.313	98	756298	54.004	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	108.000%	
43) trans-1,3-Dichloroprop...	7.619	79	117727	55.820	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	111.640%	
47) 2-Hexanone-d5	8.088	63	247914	133.631	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	133.630%#	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	408280	56.838	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	113.680%	
66) 1,2-Dichlorobenzene-d4	11.622	152	279769	64.069	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	128.140%#	
Target Compounds						Qvalue
34) Trichloroethene	5.915	95	16241	3.507	ug/L	97
46) Tetrachloroethene	7.972	164	400206	113.208	ug/L	97
51) Chlorobenzene	8.886	112	13544	1.054	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027153.D
Acq On : 02 Aug 2022 02:35
Operator : SY/MD
Sample : N3880-16 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUG9

Quant Time: Aug 02 04:10:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 02 00:27:15 2022
Response via : Initial Calibration

