

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080122\
 Data File : VW027174.D
 Acq On : 02 Aug 2022 16:32
 Operator : SY/MD
 Sample : N4005-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EY149

Quant Time: Aug 03 02:01:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 03 02:00:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	544424	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	540760	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	220434	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	189589	53.733	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 107.460%	
7) Chloroethane-d5	1.561	69	146000	55.297	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 110.600%	
11) 1,1-Dichloroethene-d2	2.101	63	237840	35.399	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 70.800%	
21) 2-Butanone-d5	3.876	46	251207	108.140	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 108.140%	
24) Chloroform-d	4.343	84	351949	49.348	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 98.700%	
26) 1,2-Dichloroethane-d4	5.027	65	207490	51.277	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 102.560%	
32) Benzene-d6	5.047	84	688319	48.103	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 96.200%	
36) 1,2-Dichloropropane-d6	6.066	67	224428	50.701	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 101.400%	
41) Toluene-d8	7.314	98	553970	42.943	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 85.880%	
43) trans-1,3-Dichloroprop...	7.619	79	90995	46.839	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 93.680%	
47) 2-Hexanone-d5	8.088	63	184123	107.742	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 107.740%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	328350	49.624	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 99.240%	
66) 1,2-Dichlorobenzene-d4	11.622	152	214906	55.473	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 110.940%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027174.D
Acq On : 02 Aug 2022 16:32
Operator : SY/MD
Sample : N4005-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EY149

Quant Time: Aug 03 02:01:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 03 02:00:39 2022
Response via : Initial Calibration

