




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
 Data File : VV027172.D
 Acq On : 02 Aug 2022 15:19
 Operator : SY/MD
 Sample : VV0801WBL04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK247

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/03/2022
 Supervised By :Mahesh Dadoda 08/03/2022

Quant Time: Aug 03 03:46:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 03 02:00:39 2022
 Response via : Initial Calibration



TIC: VV027172.D\data.ms

(7) Chloroethane-d5 (S)

1.629min (+ 0.067) 0.79 ug/L

response 2320

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	32.10	34.61
51.00	23.90	18.88
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027172.D
Acq On : 02 Aug 2022 15:19
Operator : SY/MD
Sample : VV0801WBL04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK247

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/03/2022
Supervised By :Mahesh Dadoda 08/03/2022

Quant Time: Aug 03 03:46:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 03 02:00:39 2022
Response via : Initial Calibration



TIC: VV027172.D\data.ms

(7) Chloroethane-d5 (S)

1.561min (-0.000) 57.05 ug/L m

response 167184

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	32.10	0.48#
51.00	23.90	0.26#
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
 Data File : VV027172.D
 Acq On : 02 Aug 2022 15:19
 Operator : SY/MD
 Sample : VV0801WBL04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK247

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/03/2022
 Supervised By :Mahesh Dadoda 08/03/2022

Quant Time: Aug 03 03:46:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 03 02:00:39 2022
 Response via : Initial Calibration



TIC: VV027172.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.172min (+ 0.068) 1.66 ug/L

response 12376

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	79.90	57.63
65.00	23.50	29.68
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
 Data File : VV027172.D
 Acq On : 02 Aug 2022 15:19
 Operator : SY/MD
 Sample : VV0801WBL04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK247

Manual IntegrationsAPPROVED

Quant Time: Aug 03 03:46:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 03 02:00:39 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 08/03/2022
 Supervised By :Mahesh Dadoda 08/03/2022



TIC: VV027172.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.101min (-0.003) 36.75 ug/L m

response 274046

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	79.90	2.60#
65.00	23.50	1.34#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
 Data File : VV027172.D
 Acq On : 02 Aug 2022 15:19
 Operator : SY/MD
 Sample : VV0801WBL04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK247

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/03/2022
 Supervised By :Mahesh Dadoda 08/03/2022

Quant Time: Aug 03 03:46:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 03 02:00:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	604234	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	592572	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	250501	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	211052	53.895	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	107.800%	
7) Chloroethane-d5	1.561	69	167184m	57.052	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	114.100%	
11) 1,1-Dichloroethene-d2	2.101	63	274046m	36.751	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	73.500%	
21) 2-Butanone-d5	3.873	46	269052	104.357	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	104.360%	
24) Chloroform-d	4.339	84	378742	47.848	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	95.700%	
26) 1,2-Dichloroethane-d4	5.024	65	231413	51.528	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	103.060%	
32) Benzene-d6	5.043	84	776389	49.514	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	99.020%	
36) 1,2-Dichloropropane-d6	6.063	67	251329	51.814	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	103.620%	
41) Toluene-d8	7.313	98	644134	45.566	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	91.140%	
43) trans-1,3-Dichloroprop...	7.619	79	103753	48.736	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	97.480%	
47) 2-Hexanone-d5	8.085	63	208535	111.358	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	111.360%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	366550	50.553	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	101.100%	
66) 1,2-Dichlorobenzene-d4	11.622	152	250579	56.918	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	113.840%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027172.D
Acq On : 02 Aug 2022 15:19
Operator : SY/MD
Sample : VV0801WBL04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK247

Manual IntegrationsAPPROVED

Quant Time: Aug 03 03:46:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 03 02:00:39 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 08/03/2022
Supervised By :Mahesh Dadoda 08/03/2022

