

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
 Data File : VV027128.D
 Acq On : 01 Aug 2022 11:56
 Operator : SY/MD
 Sample : VV0801MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK243

Quant Time: Aug 02 00:19:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 02 00:15:03 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.609	114	550506	50.000	ug/L	0.00
28) Chlorobenzene-d5		8.847	117	551312	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.246	152	244742	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.301	65	215478	60.396	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	120.800%	
7) Chloroethane-d5		1.558	69	125770	47.108	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.220%	
11) 1,1-Dichloroethene-d2		2.095	63	298312	43.909	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.820%	
21) 2-Butanone-d5		3.880	46	315370	134.261	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	134.260%#	
24) Chloroform-d		4.336	84	433184	60.067	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	120.140%	
26) 1,2-Dichloroethane-d4		5.021	65	267450	65.365	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	130.720%#	
32) Benzene-d6		5.040	84	890896	61.068	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	122.140%	
36) 1,2-Dichloropropane-d6		6.059	67	285386	63.238	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	126.480%#	
41) Toluene-d8		7.310	98	720213	54.761	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.520%	
43) trans-1,3-Dichloroprop...		7.616	79	117153	59.149	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	118.300%	
47) 2-Hexanone-d5		8.088	63	227401	130.520	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	130.520%#	
56) 1,1,2,2-Tetrachloroeth...		10.214	84	418798	62.082	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	124.160%#	
66) 1,2-Dichlorobenzene-d4		11.622	152	286927	66.708	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	133.420%#	
Target Compounds							Qvalue
10) 1,1,2-Trichloro-1,2,2-...		2.095	101	2888	0.735	ug/L #	23
48) 2-Hexanone		8.088	43	25746	6.312	ug/L #	63
71) Naphthalene		13.506	128	19818	1.368	ug/L #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027128.D
Acq On : 01 Aug 2022 11:56
Operator : SY/MD
Sample : VV0801MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK243

Quant Time: Aug 02 00:19:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 02 00:15:03 2022
Response via : Initial Calibration

