

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
 Data File : VV027129.D
 Acq On : 01 Aug 2022 14:52
 Operator : SY/MD
 Sample : VV0801WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK244

Quant Time: Aug 02 00:19:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 02 00:15:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	528894	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	542799	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	237692	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	194283	56.680	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	113.360%	
7) Chloroethane-d5	1.561	69	156162	60.882	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	121.760%	
11) 1,1-Dichloroethene-d2	2.102	63	279492	42.820	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	85.640%	
21) 2-Butanone-d5	3.876	46	246524	109.240	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	109.240%	
24) Chloroform-d	4.343	84	388117	56.017	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	112.040%	
26) 1,2-Dichloroethane-d4	5.027	65	229947	58.495	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	117.000%	
32) Benzene-d6	5.044	84	790959	55.068	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	110.140%	
36) 1,2-Dichloropropane-d6	6.063	67	250386	56.353	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	112.700%	
41) Toluene-d8	7.314	98	663676	51.254	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	102.500%	
43) trans-1,3-Dichloroprop...	7.619	79	97820	50.163	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	100.320%	
47) 2-Hexanone-d5	8.088	63	166804	97.241	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	97.240%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	346393	52.154	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	104.300%	
66) 1,2-Dichlorobenzene-d4	11.622	152	245395	58.744	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	117.480%	
Target Compounds						
10) 1,1,2-Trichloro-1,2,2-...	2.102	101	2794	0.740	ug/L #	20
61) 1,2,3-Trichloropropane	11.246	75	26248	5.319	ug/L #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027129.D
Acq On : 01 Aug 2022 14:52
Operator : SY/MD
Sample : VV0801WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK244

Quant Time: Aug 02 00:19:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 02 00:15:03 2022
Response via : Initial Calibration

