

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
 Data File : VV027156.D
 Acq On : 02 Aug 2022 05:01
 Operator : SY/MD
 Sample : VV0801WBL03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK246

Quant Time: Aug 02 05:38:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 02 04:33:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	595735	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	564411	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	231667	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	189215	49.008	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.020%
7) Chloroethane-d5	1.561	69	160792	55.654	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.300%
11) 1,1-Dichloroethene-d2	2.105	63	257666	35.047	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.100%
21) 2-Butanone-d5	3.876	46	259993	102.282	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.280%
24) Chloroform-d	4.343	84	379124	48.580	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.160%
26) 1,2-Dichloroethane-d4	5.027	65	231324	52.243	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.480%
32) Benzene-d6	5.047	84	779372	52.184	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.360%
36) 1,2-Dichloropropane-d6	6.066	67	249966	54.104	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.200%
41) Toluene-d8	7.313	98	631757	46.921	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.840%
43) trans-1,3-Dichloroprop...	7.619	79	93001	45.865	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.740%
47) 2-Hexanone-d5	8.088	63	184239	103.293	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.290%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	346980	50.242	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.480%
66) 1,2-Dichlorobenzene-d4	11.622	152	230392	56.587	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080122\
Data File : VV027156.D
Acq On : 02 Aug 2022 05:01
Operator : SY/MD
Sample : VV0801WBL03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK246

Quant Time: Aug 02 05:38:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 02 04:33:39 2022
Response via : Initial Calibration

