

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080524\
 Data File : VV036781.D
 Acq On : 05 Aug 2024 13:08
 Operator : SY/MD
 Sample : P3437-03
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E05W2

Quant Time: Aug 06 03:09:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 06 02:51:35 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	370561	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	348697	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	170719	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	206137	69.648	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	139.300%#	
7) Chloroethane-d5	1.532	69	162686	69.701	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	139.400%#	
11) 1,1-Dichloroethene-d2	2.060	65	88556	69.948	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	139.900%#	
21) 2-Butanone-d5	3.796	46	200124	127.617	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	127.620%	
24) Chloroform-d	4.246	84	353735	66.600	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	133.200%#	
26) 1,2-Dichloroethane-d4	4.937	65	237072	69.422	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	138.840%#	
32) Benzene-d6	4.953	84	746242	70.855	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	141.720%#	
36) 1,2-Dichloropropane-d6	5.985	67	235486	69.306	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	138.620%#	
41) Toluene-d8	7.239	98	681785	70.571	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	141.140%#	
43) trans-1,3-Dichloroprop...	7.551	79	113410	64.135	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	128.280%#	
47) 2-Hexanone-d5	8.024	63	112479	115.133	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	115.130%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	246517	63.188	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	126.380%#	
66) 1,2-Dichlorobenzene-d4	11.554	152	239720	68.909	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	137.820%#	
Target Compounds						Qvalue
13) Acetone	2.114	43	8632	6.579	ug/L	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080524\
Data File : VV036781.D
Acq On : 05 Aug 2024 13:08
Operator : SY/MD
Sample : P3437-03
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E05W2

Quant Time: Aug 06 03:09:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 06 02:51:35 2024
Response via : Initial Calibration

