

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080624\
 Data File : VV036806.D
 Acq On : 06 Aug 2024 12:05
 Operator : SY/MD
 Sample : P3437-05
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E05W5

Quant Time: Aug 07 05:01:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 07 04:57:18 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.532	114	219122	50.000	ug/L	0.00
28) Chlorobenzene-d5		8.783	117	203252	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.182	152	101310	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	117867	67.347	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	134.700%	
7) Chloroethane-d5		1.532	69	94963	68.805	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	137.600%#	
11) 1,1-Dichloroethene-d2		2.060	65	50900	67.991	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	135.980%#	
21) 2-Butanone-d5		3.799	46	128766	138.863	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	138.860%#	
24) Chloroform-d		4.246	84	219153	69.778	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	139.560%#	
26) 1,2-Dichloroethane-d4		4.941	65	148005	73.294	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	146.580%#	
32) Benzene-d6		4.953	84	452358	73.687	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	147.380%#	
36) 1,2-Dichloropropane-d6		5.986	67	147048	74.247	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	148.500%#	
41) Toluene-d8		7.240	98	402484	71.472	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	142.940%#	
43) trans-1,3-Dichloroprop...		7.555	79	69950	67.865	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	135.740%#	
47) 2-Hexanone-d5		8.027	63	72583	127.460	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	127.460%	
56) 1,1,2,2-Tetrachloroeth...		10.149	84	160542	70.598	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	141.200%#	
66) 1,2-Dichlorobenzene-d4		11.558	152	151062	73.174	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	146.340%#	
Target Compounds							Qvalue
13) Acetone		2.114	43	7842	10.108	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080624\
Data File : VV036806.D
Acq On : 06 Aug 2024 12:05
Operator : SY/MD
Sample : P3437-05
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E05W5

Quant Time: Aug 07 05:01:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 07 04:57:18 2024
Response via : Initial Calibration

