

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080624\
 Data File : VW036820.D
 Acq On : 06 Aug 2024 17:55
 Operator : SY/MD
 Sample : P3480-01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AA3

Quant Time: Aug 07 05:05:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 07 04:57:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	435464	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	407671	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	201390	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	136602	39.275	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	78.540%		
7) Chloroethane-d5	1.529	69	118898	43.348	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	86.700%		
11) 1,1-Dichloroethene-d2	2.057	65	62056	41.711	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	83.420%		
21) 2-Butanone-d5	3.796	46	206034	111.804	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	111.800%		
24) Chloroform-d	4.243	84	290373	46.522	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.040%		
26) 1,2-Dichloroethane-d4	4.937	65	195742	48.776	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.560%		
32) Benzene-d6	4.953	84	586676	47.646	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.300%		
36) 1,2-Dichloropropane-d6	5.986	67	193357	48.675	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	97.340%		
41) Toluene-d8	7.240	98	514874	45.584	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.160%		
43) trans-1,3-Dichloroprop...	7.551	79	90841	43.941	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	87.880%		
47) 2-Hexanone-d5	8.024	63	121504	106.379	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	106.380%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	241863	53.027	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	106.060%		
66) 1,2-Dichlorobenzene-d4	11.558	152	201761	49.165	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.320%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080624\
Data File : VV036820.D
Acq On : 06 Aug 2024 17:55
Operator : SY/MD
Sample : P3480-01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AA3

Quant Time: Aug 07 05:05:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 07 04:57:18 2024
Response via : Initial Calibration

