

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080724\
 Data File : VV036826.D
 Acq On : 07 Aug 2024 09:34
 Operator : SY/MD
 Sample : VV0807MBL01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Aug 08 02:35:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 08 02:35:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	467196	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	440489	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	214322	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	133817	35.861	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.720%
7) Chloroethane-d5	1.532	69	118221	40.174	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.340%
11) 1,1-Dichloroethene-d2	2.060	65	60497	37.901	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.800%
21) 2-Butanone-d5	3.796	46	179961	91.023	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.020%
24) Chloroform-d	4.243	84	290797	43.426	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.860%
26) 1,2-Dichloroethane-d4	4.937	65	197240	45.811	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.620%
32) Benzene-d6	4.953	84	587163	44.133	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.260%
36) 1,2-Dichloropropane-d6	5.985	67	198741	46.303	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.600%
41) Toluene-d8	7.239	98	506856	41.531	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.060%
43) trans-1,3-Dichloroprop...	7.551	79	91941	41.159	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.320%
47) 2-Hexanone-d5	8.024	63	100196	81.188	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.190%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	225609	45.778	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.560%
66) 1,2-Dichlorobenzene-d4	11.554	152	198218	45.387	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080724\
Data File : VV036826.D
Acq On : 07 Aug 2024 09:34
Operator : SY/MD
Sample : VV0807MBL01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 08 02:35:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 08 02:35:03 2024
Response via : Initial Calibration

