

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080724\
 Data File : VV036837.D
 Acq On : 07 Aug 2024 14:49
 Operator : SY/MD
 Sample : VIBLK326
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK326

Quant Time: Aug 08 02:38:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 08 02:35:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	380367	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	341049	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	155358	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	100143	32.963	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.920%
7) Chloroethane-d5	1.532	69	100920	42.124	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.240%
11) 1,1-Dichloroethene-d2	2.056	65	44450	34.205	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.400%
21) 2-Butanone-d5	3.789	46	170216	105.747	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.750%
24) Chloroform-d	4.243	84	264791	48.569	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.140%
26) 1,2-Dichloroethane-d4	4.934	65	184389	52.603	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.200%
32) Benzene-d6	4.950	84	491496	47.714	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.420%
36) 1,2-Dichloropropane-d6	5.982	67	176543	53.124	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.240%
41) Toluene-d8	7.236	98	400335	42.367	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.740%
43) trans-1,3-Dichloroprop...	7.551	79	80867	46.757	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.520%
47) 2-Hexanone-d5	8.024	63	95208	99.640	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.640%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	206524	54.124	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.240%
66) 1,2-Dichlorobenzene-d4	11.554	152	153062	48.349	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080724\
Data File : VV036837.D
Acq On : 07 Aug 2024 14:49
Operator : SY/MD
Sample : VIBLK326
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK326

Quant Time: Aug 08 02:38:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 08 02:35:03 2024
Response via : Initial Calibration

