

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080724\
 Data File : VV036841.D
 Acq On : 07 Aug 2024 16:38
 Operator : SY/MD
 Sample : P3480-09
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AD4

Quant Time: Aug 08 02:39:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 08 02:35:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	495333	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	461827	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	226822	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	147733	37.341	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	74.680%		
7) Chloroethane-d5	1.532	69	124125	39.784	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	79.560%		
11) 1,1-Dichloroethene-d2	2.060	65	63871	37.742	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	75.480%		
21) 2-Butanone-d5	3.796	46	161853	77.214	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	77.210%		
24) Chloroform-d	4.246	84	283868	39.983	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	79.960%		
26) 1,2-Dichloroethane-d4	4.937	65	190773	41.792	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.580%		
32) Benzene-d6	4.953	84	578606	41.481	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	82.960%		
36) 1,2-Dichloropropane-d6	5.986	67	192265	42.724	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	85.440%		
41) Toluene-d8	7.239	98	502780	39.294	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	78.580%#		
43) trans-1,3-Dichloroprop...	7.551	79	87795	37.487	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.980%		
47) 2-Hexanone-d5	8.024	63	85752	66.274	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	66.270%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	191521	37.066	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	74.140%		
66) 1,2-Dichlorobenzene-d4	11.554	152	198816	43.015	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	86.020%		
Target Compounds					Qvalue	
13) Acetone	2.114	43	31218	17.800	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080724\
Data File : VV036841.D
Acq On : 07 Aug 2024 16:38
Operator : SY/MD
Sample : P3480-09
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AD4

Quant Time: Aug 08 02:39:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 08 02:35:03 2024
Response via : Initial Calibration

