

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080822\
 Data File : VV027245.D
 Acq On : 08 Aug 2022 13:55
 Operator : SY/MD
 Sample : VIBLK272
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK272

Quant Time: Aug 09 02:30:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 09 02:29:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	446512	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	452811	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	220862	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	108112	35.514	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	71.020%	
7) Chloroethane-d5	1.558	69	94384	42.410	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	84.820%	
11) 1,1-Dichloroethene-d2	2.098	63	150691	27.062	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	54.120%#	
21) 2-Butanone-d5	3.863	46	255792	94.398	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	94.400%	
24) Chloroform-d	4.336	84	272715	43.734	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.460%	
26) 1,2-Dichloroethane-d4	5.021	65	202755	52.370	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.740%	
32) Benzene-d6	5.040	84	582132	53.610	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	107.220%	
36) 1,2-Dichloropropane-d6	6.063	67	200676	51.953	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.900%	
41) Toluene-d8	7.310	98	488513	52.780	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	105.560%	
43) trans-1,3-Dichloroprop...	7.616	79	87035	51.695	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	103.380%	
47) 2-Hexanone-d5	8.082	63	187180	100.613	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	100.610%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	287054	45.603	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	91.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	205793	53.956	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	107.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080822\
Data File : VV027245.D
Acq On : 08 Aug 2022 13:55
Operator : SY/MD
Sample : VIBLK272
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK272

Quant Time: Aug 09 02:30:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 09 02:29:12 2022
Response via : Initial Calibration

