

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080824\
 Data File : VV036853.D
 Acq On : 08 Aug 2024 15:41
 Operator : SY/MD
 Sample : P3480-17
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AB8

Quant Time: Aug 09 02:43:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 09 02:41:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	448543	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	417871	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	202248	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	104829	29.261	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	58.520%#
7) Chloroethane-d5	1.532	69	97377	34.467	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	68.940%#
11) 1,1-Dichloroethene-d2	2.060	65	47906	31.261	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.520%
21) 2-Butanone-d5	3.796	46	161891	85.288	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.290%
24) Chloroform-d	4.243	84	251664	39.145	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.280%
26) 1,2-Dichloroethane-d4	4.937	65	166855	40.366	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.740%
32) Benzene-d6	4.953	84	477734	37.852	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.700%
36) 1,2-Dichloropropane-d6	5.985	67	167090	41.036	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.080%
41) Toluene-d8	7.239	98	408984	35.325	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	70.660%#
43) trans-1,3-Dichloroprop...	7.551	79	72479	34.203	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.400%
47) 2-Hexanone-d5	8.024	63	88437	75.538	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.540%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	204910	43.829	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.660%
66) 1,2-Dichlorobenzene-d4	11.554	152	172781	41.924	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080824\
Data File : VV036853.D
Acq On : 08 Aug 2024 15:41
Operator : SY/MD
Sample : P3480-17
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AB8

Quant Time: Aug 09 02:43:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 09 02:41:02 2024
Response via : Initial Calibration

