

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
 Data File : VV036871.D
 Acq On : 09 Aug 2024 12:37
 Operator : SY/MD
 Sample : P3480-21RE
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AD5RE

Quant Time: Aug 10 01:17:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	465150	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	449937	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	213544	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	277015	74.562	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 149.120%#		
7) Chloroethane-d5	1.532	69	212990	72.697	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 145.400%#		
11) 1,1-Dichloroethene-d2	2.056	65	117302	73.813	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 147.620%#		
21) 2-Butanone-d5	3.799	46	252028	128.034	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 128.030%		
24) Chloroform-d	4.243	84	451280	67.688	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 135.380%#		
26) 1,2-Dichloroethane-d4	4.937	65	299785	69.935	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 139.860%#		
32) Benzene-d6	4.953	84	962105	70.796	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 141.600%#		
36) 1,2-Dichloropropane-d6	5.985	67	308167	70.289	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 140.580%#		
41) Toluene-d8	7.239	98	860636	69.039	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 138.080%#		
43) trans-1,3-Dichloroprop...	7.551	79	138007	60.485	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 120.960%		
47) 2-Hexanone-d5	8.027	63	127338	101.014	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 101.010%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	290652	57.738	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 115.480%		
66) 1,2-Dichlorobenzene-d4	11.554	152	302844	69.596	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 139.200%#		
Target Compounds						
13) Acetone	2.114	43	13848	8.408	ug/L	98
42) Toluene	7.333	91	13571	0.947	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
Data File : VV036871.D
Acq On : 09 Aug 2024 12:37
Operator : SY/MD
Sample : P3480-21RE
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AD5RE

Quant Time: Aug 10 01:17:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 01:11:12 2024
Response via : Initial Calibration

