

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
 Data File : VV036873.D
 Acq On : 09 Aug 2024 13:30
 Operator : SY/MD
 Sample : P3533-01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AC4

Quant Time: Aug 10 01:17:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	491971	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	468831	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	231753	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	304487	77.489	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 154.980%#		
7) Chloroethane-d5	1.532	69	232153	74.918	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 149.840%#		
11) 1,1-Dichloroethene-d2	2.057	65	126399	75.201	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 150.400%#		
21) 2-Butanone-d5	3.796	46	270200	129.783	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 129.780%		
24) Chloroform-d	4.240	84	501654	71.141	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 142.280%#		
26) 1,2-Dichloroethane-d4	4.937	65	325820	71.865	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 143.720%#		
32) Benzene-d6	4.950	84	1053258	74.381	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 148.760%#		
36) 1,2-Dichloropropane-d6	5.982	67	336173	73.587	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 147.180%#		
41) Toluene-d8	7.236	98	944657	72.725	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 145.440%#		
43) trans-1,3-Dichloroprop...	7.551	79	153837	64.705	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 129.420%#		
47) 2-Hexanone-d5	8.024	63	146030	111.173	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 111.170%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	326806	62.303	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 124.600%#		
66) 1,2-Dichlorobenzene-d4	11.555	152	339609	71.913	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 143.820%#		
Target Compounds						
13) Acetone	2.114	43	38817	22.284	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
Data File : VV036873.D
Acq On : 09 Aug 2024 13:30
Operator : SY/MD
Sample : P3533-01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AC4

Quant Time: Aug 10 01:17:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 01:11:12 2024
Response via : Initial Calibration

