

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
 Data File : VV036879.D
 Acq On : 09 Aug 2024 16:22
 Operator : SY/MD
 Sample : P3480-21
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AD5

Quant Time: Aug 10 01:19:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	705549	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	661958	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	321225	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	267294	47.432	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.860%
7) Chloroethane-d5	1.532	69	208455	46.907	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.820%
11) 1,1-Dichloroethene-d2	2.060	65	114802	47.626	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.260%
21) 2-Butanone-d5	3.799	46	253607	84.938	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.940%
24) Chloroform-d	4.246	84	450945	44.592	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.180%
26) 1,2-Dichloroethane-d4	4.937	65	298591	45.923	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.840%
32) Benzene-d6	4.953	84	952004	47.616	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.240%
36) 1,2-Dichloropropane-d6	5.985	67	303211	47.008	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.020%
41) Toluene-d8	7.239	98	848119	46.243	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.480%
43) trans-1,3-Dichloroprop...	7.551	79	134121	39.954	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.900%
47) 2-Hexanone-d5	8.027	63	144969	78.166	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.170%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	316971	42.798	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.600%
66) 1,2-Dichlorobenzene-d4	11.558	152	305942	46.739	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.480%
Target Compounds						
13) Acetone	2.117	43	14319	5.732	ug/L	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
Data File : VV036879.D
Acq On : 09 Aug 2024 16:22
Operator : SY/MD
Sample : P3480-21
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AD5

Quant Time: Aug 10 01:19:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 01:11:12 2024
Response via : Initial Calibration

