

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
 Data File : VV036866.D
 Acq On : 09 Aug 2024 10:46
 Operator : SY/MD
 Sample : VV0809MBL01
 Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK365

Quant Time: Aug 10 01:15:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	444415	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	420067	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	202456	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	138527	39.026	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.060%
7) Chloroethane-d5	1.532	69	120399	43.012	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.020%
11) 1,1-Dichloroethene-d2	2.059	65	61488	40.497	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.000%
21) 2-Butanone-d5	3.796	46	183456	97.547	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.550%
24) Chloroform-d	4.243	84	292542	45.926	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.860%
26) 1,2-Dichloroethane-d4	4.937	65	192858	47.090	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.180%
32) Benzene-d6	4.953	84	579245	45.655	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.300%
36) 1,2-Dichloropropane-d6	5.985	67	196675	48.049	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.100%
41) Toluene-d8	7.239	98	500383	42.994	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.980%
43) trans-1,3-Dichloroprop...	7.551	79	86159	40.446	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.900%
47) 2-Hexanone-d5	8.024	63	101128	85.927	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.930%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	227374	48.379	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.760%
66) 1,2-Dichlorobenzene-d4	11.554	152	201792	48.913	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.820%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
Data File : VV036866.D
Acq On : 09 Aug 2024 10:46
Operator : SY/MD
Sample : VV0809MBL01
Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK365

Quant Time: Aug 10 01:15:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 01:11:12 2024
Response via : Initial Calibration

