

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
 Data File : VV036867.D
 Acq On : 09 Aug 2024 11:08
 Operator : SY/MD
 Sample : VV0809WBL01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK366

Quant Time: Aug 10 01:15:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	439077	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	409364	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	198408	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	140584	40.087	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.180%
7) Chloroethane-d5	1.532	69	120100	43.426	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.860%
11) 1,1-Dichloroethene-d2	2.060	65	59759	39.836	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.680%
21) 2-Butanone-d5	3.796	46	172705	92.947	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.950%
24) Chloroform-d	4.243	84	286237	45.482	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.960%
26) 1,2-Dichloroethane-d4	4.937	65	187323	46.294	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.580%
32) Benzene-d6	4.953	84	566905	45.850	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.700%
36) 1,2-Dichloropropane-d6	5.985	67	189610	47.534	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.060%
41) Toluene-d8	7.239	98	484199	42.691	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.380%
43) trans-1,3-Dichloroprop...	7.551	79	85524	41.198	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.400%
47) 2-Hexanone-d5	8.024	63	97983	85.431	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.430%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	218346	47.673	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.340%
66) 1,2-Dichlorobenzene-d4	11.554	152	196665	48.643	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080924\
Data File : VV036867.D
Acq On : 09 Aug 2024 11:08
Operator : SY/MD
Sample : VV0809WBL01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK366

Quant Time: Aug 10 01:15:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 01:11:12 2024
Response via : Initial Calibration

