

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081322\
 Data File : VV027377.D
 Acq On : 13 Aug 2022 18:50
 Operator : SY/MD
 Sample : N4133-04DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E45DL

Quant Time: Aug 13 21:54:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 13 21:49:38 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	140878	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	147367	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62864	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	57598	4.706	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	94.200%	
7) Chloroethane-d5	1.532	69	36460	4.102	ug/L	-0.03
Spiked Amount	5.000	Range 65 - 130	Recovery	=	82.000%	
11) 1,1-Dichloroethene-d2	2.092	63	76648	3.453	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.000%	
20) 2-Butanone-d5	3.870	46	118790	57.308	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.620%	
24) Chloroform-d	4.336	84	106928	4.764	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.200%	
26) 1,2-Dichloroethane-d4	5.024	65	54926	5.136	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.800%	
32) Benzene-d6	5.040	84	203705	4.686	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.800%	
36) 1,2-Dichloropropane-d6	6.063	67	68943	5.129	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.600%	
41) Toluene-d8	7.313	98	169329	4.255	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.000%	
43) trans-1,3-Dichloroprop...	7.622	79	21376	4.591	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.800%	
46) 2-Hexanone-d5	8.085	63	81802	45.400	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.800%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	46027	4.733	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	58603	5.401	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.000%	
Target Compounds						
5) Vinyl chloride	1.304	62	88037	5.636	ug/L	98
13) Acetone	2.156	43	7444	4.860	ug/L	100
16) Methylene chloride	2.494	84	5796	0.356	ug/L	96
18) trans-1,2-Dichloroethene	2.751	96	6103	0.489	ug/L	92
22) cis-1,2-Dichloroethene	3.905	96	4283	0.349	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081322\
Data File : VV027377.D
Acq On : 13 Aug 2022 18:50
Operator : SY/MD
Sample : N4133-04DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5E45DL

Quant Time: Aug 13 21:54:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 13 21:49:38 2022
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 5.636 ug/L
 RT: 1.304 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VV027377.D
 Acq: 13 Aug 2022 18:50




Instrument :
 MSVOA_V
 ClientSampleId :
 F5E45DL


Tgt Ion: 62 Resp: 88037
 Ion Ratio Lower Upper
 62 100
 64 35.3 23.7 44.1



#13
 Acetone
 Concen: 4.860 ug/L
 RT: 2.156 min Scan# 347
 Delta R.T. -0.013 min
 Lab File: VV027377.D
 Acq: 13 Aug 2022 18:50

Tgt Ion: 43 Resp: 7444
 Ion Ratio Lower Upper
 43 100
 58 31.4 0.0 63.2









#16
Methylene chloride
Concen: 0.356 ug/L
RT: 2.494 min Scan# 41
Delta R.T. -0.003 min
Lab File: VV027377.D
Acq: 13 Aug 2022 18:50




Instrument :
MSVOA_V
ClientSampleId :
F5E45DL


Tgt Ion: 84 Resp: 5796
Ion Ratio Lower Upper
84 100
86 64.4 45.1 83.9
49 123.3 91.2 169.4



#18
trans-1,2-Dichloroethene
Concen: 0.489 ug/L
RT: 2.751 min Scan# 532
Delta R.T. -0.000 min
Lab File: VV027377.D
Acq: 13 Aug 2022 18:50

Tgt Ion: 96 Resp: 6103
Ion Ratio Lower Upper
96 100
61 131.0 99.6 185.0
98 59.4 45.1 83.7





#22
 cis-1,2-Dichloroethene
 Concen: 0.349 ug/L
 RT: 3.905 min Scan# 891
 Delta R.T. 0.003 min
 Lab File: VV027377.D
 Acq: 13 Aug 2022 18:50

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E45DL

Tgt Ion: 96 Resp: 4283
 Ion Ratio Lower Upper
 96 100
 61 122.8 91.3 169.5
 68 0.0 0.0 0.0

