

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081324\
 Data File : VV036916.D
 Acq On : 13 Aug 2024 13:34
 Operator : SY/MD
 Sample : P3533-05
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Aug 14 04:11:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 14 04:10:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	716064	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	666369	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	324698	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	259360	45.348	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.700%
7) Chloroethane-d5	1.529	69	208034	46.125	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.240%
11) 1,1-Dichloroethene-d2	2.056	65	112688	46.062	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.120%
21) 2-Butanone-d5	3.789	46	235552	77.733	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.730%
24) Chloroform-d	4.240	84	452076	44.047	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.100%
26) 1,2-Dichloroethane-d4	4.931	65	294289	44.596	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.200%
32) Benzene-d6	4.947	84	931020	46.258	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.520%
36) 1,2-Dichloropropane-d6	5.979	67	302177	46.537	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.080%
41) Toluene-d8	7.236	98	815221	44.155	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.320%
43) trans-1,3-Dichloroprop...	7.548	79	126100	37.316	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.640%
47) 2-Hexanone-d5	8.024	63	129049	69.122	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	69.120%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	318488	42.718	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.440%
66) 1,2-Dichlorobenzene-d4	11.554	152	298914	45.177	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081324\
Data File : VV036916.D
Acq On : 13 Aug 2024 13:34
Operator : SY/MD
Sample : P3533-05
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Aug 14 04:11:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 14 04:10:20 2024
Response via : Initial Calibration

