

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081524\
 Data File : VV036940.D
 Acq On : 15 Aug 2024 15:20
 Operator : SY/MD
 Sample : P3603-01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0CL1

Quant Time: Aug 16 02:16:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 16 02:14:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	648451	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	623292	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	297223	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	250866	48.437	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.880%
7) Chloroethane-d5	1.532	69	198585	48.621	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.240%
11) 1,1-Dichloroethene-d2	2.060	65	107657	48.594	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.180%
21) 2-Butanone-d5	3.796	46	244471	89.088	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.090%
24) Chloroform-d	4.243	84	446146	48.002	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	4.937	65	285031	47.697	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.400%
32) Benzene-d6	4.953	84	921221	48.934	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.860%
36) 1,2-Dichloropropane-d6	5.986	67	292211	48.113	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.220%
41) Toluene-d8	7.239	98	817304	47.328	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.660%
43) trans-1,3-Dichloroprop...	7.551	79	127273	40.266	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.540%
47) 2-Hexanone-d5	8.027	63	138026	79.040	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.040%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	330469	47.389	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.780%
66) 1,2-Dichlorobenzene-d4	11.554	152	301609	49.798	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.600%
Target Compounds						
					Qvalue	
46) Tetrachloroethene	7.902	164	15742	4.442	ug/L	96
49) Dibromochloromethane	8.175	129	6056	1.232	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081524\
Data File : VV036940.D
Acq On : 15 Aug 2024 15:20
Operator : SY/MD
Sample : P3603-01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0CL1

Quant Time: Aug 16 02:16:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 16 02:14:55 2024
Response via : Initial Calibration

