

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081624\
 Data File : VV036947.D
 Acq On : 16 Aug 2024 09:26
 Operator : SY/MD
 Sample : VV0816MBL01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Aug 16 23:06:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 16 23:06:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	630473	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	604694	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	293569	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	235890	46.844	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	93.680%		
7) Chloroethane-d5	1.532	69	195229	49.162	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	98.320%		
11) 1,1-Dichloroethene-d2	2.060	65	105424	48.943	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	97.880%		
21) 2-Butanone-d5	3.796	46	224697	84.217	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	84.220%		
24) Chloroform-d	4.243	84	441416	48.847	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.700%		
26) 1,2-Dichloroethane-d4	4.937	65	286568	49.322	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.640%		
32) Benzene-d6	4.953	84	903322	49.459	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.920%		
36) 1,2-Dichloropropane-d6	5.986	67	292096	49.573	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.140%		
41) Toluene-d8	7.240	98	811255	48.422	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.840%		
43) trans-1,3-Dichloroprop...	7.551	79	126651	41.302	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	82.600%		
47) 2-Hexanone-d5	8.024	63	125769	74.236	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	74.240%		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	315244	46.596	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	93.200%		
66) 1,2-Dichlorobenzene-d4	11.555	152	292187	48.843	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.680%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081624\
Data File : VV036947.D
Acq On : 16 Aug 2024 09:26
Operator : SY/MD
Sample : VV0816MBL01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 16 23:06:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 16 23:06:03 2024
Response via : Initial Calibration

