

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081624\
 Data File : VW036957.D
 Acq On : 16 Aug 2024 15:57
 Operator : SY/MD
 Sample : P3648-12
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JAAS8

Quant Time: Aug 16 23:09:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 16 23:06:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	610779	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	580678	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	278511	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	231978	47.552	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.100%
7) Chloroethane-d5	1.532	69	191163	49.690	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.380%
11) 1,1-Dichloroethene-d2	2.056	65	102230	48.991	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.980%
21) 2-Butanone-d5	3.799	46	254593	98.499	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.500%
24) Chloroform-d	4.243	84	427852	48.873	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.740%
26) 1,2-Dichloroethane-d4	4.937	65	278593	49.495	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.000%
32) Benzene-d6	4.953	84	879380	50.140	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.280%
36) 1,2-Dichloropropane-d6	5.985	67	280977	49.658	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.320%
41) Toluene-d8	7.239	98	778593	48.395	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.780%
43) trans-1,3-Dichloroprop...	7.551	79	118434	40.219	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.440%
47) 2-Hexanone-d5	8.027	63	139925	86.007	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.010%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	323118	49.735	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.480%
66) 1,2-Dichlorobenzene-d4	11.557	152	285566	50.317	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081624\
Data File : VV036957.D
Acq On : 16 Aug 2024 15:57
Operator : SY/MD
Sample : P3648-12
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JAAS8

Quant Time: Aug 16 23:09:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 16 23:06:03 2024
Response via : Initial Calibration

