

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081624\
 Data File : VW036959.D
 Acq On : 16 Aug 2024 16:41
 Operator : SY/MD
 Sample : P3648-14
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JAATO

Quant Time: Aug 16 23:09:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 16 23:06:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	590593	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	572707	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	275466	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	224700	47.635	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 95.260%			
7) Chloroethane-d5	1.529	69	187036	50.279	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 100.560%			
11) 1,1-Dichloroethene-d2	2.056	65	102087	50.594	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 101.180%			
21) 2-Butanone-d5	3.796	46	251996	100.827	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 100.830%			
24) Chloroform-d	4.246	84	456969	53.983	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 107.960%			
26) 1,2-Dichloroethane-d4	4.937	65	279985	51.443	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 102.880%			
32) Benzene-d6	4.953	84	865687	50.046	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 100.100%			
36) 1,2-Dichloropropane-d6	5.985	67	280389	50.244	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 100.480%			
41) Toluene-d8	7.239	98	771320	48.610	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 97.220%			
43) trans-1,3-Dichloroprop...	7.551	79	118008	40.633	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 81.260%			
47) 2-Hexanone-d5	8.027	63	143373	89.353	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 89.350%			
56) 1,1,2,2-Tetrachloroeth...	10.149	84	318591	49.721	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 99.440%			
66) 1,2-Dichlorobenzene-d4	11.554	152	280882	50.039	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 100.080%			
Target Compounds					Qvalue	
25) Chloroform	4.268	83	725980	90.533	ug/L	97
38) Bromodichloromethane	6.435	83	47518	7.604	ug/L #	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081624\
Data File : VV036959.D
Acq On : 16 Aug 2024 16:41
Operator : SY/MD
Sample : P3648-14
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JAATO

Quant Time: Aug 16 23:09:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 16 23:06:03 2024
Response via : Initial Calibration

