

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081624\
 Data File : VW036961.D
 Acq On : 16 Aug 2024 17:26
 Operator : SY/MD
 Sample : P3648-07
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JAAS5

Quant Time: Aug 16 23:10:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 16 23:06:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	627124	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	598557	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	293377	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	220115	43.945	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.880%
7) Chloroethane-d5	1.529	69	183279	46.399	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.800%
11) 1,1-Dichloroethene-d2	2.057	65	101349	47.302	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.600%
21) 2-Butanone-d5	3.799	46	236496	89.113	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.110%
24) Chloroform-d	4.243	84	409507	45.558	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.120%
26) 1,2-Dichloroethane-d4	4.937	65	272237	47.105	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.220%
32) Benzene-d6	4.953	84	852025	47.129	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.260%
36) 1,2-Dichloropropane-d6	5.986	67	276550	47.416	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.840%
41) Toluene-d8	7.240	98	750309	45.244	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.480%
43) trans-1,3-Dichloroprop...	7.551	79	112219	36.971	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.940%
47) 2-Hexanone-d5	8.027	63	142292	84.850	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.850%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	316404	47.247	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.500%
66) 1,2-Dichlorobenzene-d4	11.558	152	274985	45.998	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081624\
Data File : VV036961.D
Acq On : 16 Aug 2024 17:26
Operator : SY/MD
Sample : P3648-07
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JAAS5

Quant Time: Aug 16 23:10:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 16 23:06:03 2024
Response via : Initial Calibration

