

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081624\
 Data File : VW036954.D
 Acq On : 16 Aug 2024 14:51
 Operator : SY/MD
 Sample : P3648-06
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JAAS4

Quant Time: Aug 16 23:08:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 16 23:06:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	611815	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	595099	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	281497	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	226365	46.323	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	92.640%	
7) Chloroethane-d5	1.532	69	188054	48.799	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	97.600%	
11) 1,1-Dichloroethene-d2	2.056	65	102163	48.876	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	97.760%	
21) 2-Butanone-d5	3.799	46	243132	93.906	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	93.910%	
24) Chloroform-d	4.246	84	422212	48.147	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.300%	
26) 1,2-Dichloroethane-d4	4.937	65	276202	48.987	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.980%	
32) Benzene-d6	4.953	84	880154	48.968	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.940%	
36) 1,2-Dichloropropane-d6	5.985	67	285627	49.257	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	98.520%	
41) Toluene-d8	7.239	98	774228	46.957	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.920%	
43) trans-1,3-Dichloroprop...	7.551	79	122895	40.723	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	81.440%	
47) 2-Hexanone-d5	8.027	63	139909	83.913	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	83.910%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	325015	48.815	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	97.620%	
66) 1,2-Dichlorobenzene-d4	11.554	152	283734	49.464	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	98.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081624\
Data File : VV036954.D
Acq On : 16 Aug 2024 14:51
Operator : SY/MD
Sample : P3648-06
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JAAS4

Quant Time: Aug 16 23:08:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 16 23:06:03 2024
Response via : Initial Calibration

