

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081924\
 Data File : VV036967.D
 Acq On : 19 Aug 2024 09:56
 Operator : SY/MD
 Sample : VV0819MBL01
 Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK377

Quant Time: Aug 20 04:43:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 20 04:42:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	668168	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	637783	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	304836	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	212034	39.731	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.460%
7) Chloroethane-d5	1.532	69	181440	43.112	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.220%
11) 1,1-Dichloroethene-d2	2.057	65	93097	40.782	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.560%
21) 2-Butanone-d5	3.799	46	237043	83.832	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.830%
24) Chloroform-d	4.243	84	420443	43.901	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	4.937	65	276473	44.900	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.800%
32) Benzene-d6	4.953	84	840971	43.656	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.320%
36) 1,2-Dichloropropane-d6	5.986	67	284811	45.829	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.660%
41) Toluene-d8	7.240	98	727047	41.145	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.280%
43) trans-1,3-Dichloroprop...	7.551	79	121124	37.450	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.900%
47) 2-Hexanone-d5	8.024	63	139678	78.168	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.170%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	328905	46.093	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.180%
66) 1,2-Dichlorobenzene-d4	11.555	152	283168	45.586	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081924\
Data File : VV036967.D
Acq On : 19 Aug 2024 09:56
Operator : SY/MD
Sample : VV0819MBL01
Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK377

Quant Time: Aug 20 04:43:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 20 04:42:25 2024
Response via : Initial Calibration

