

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082724\
 Data File : VW037103.D
 Acq On : 27 Aug 2024 21:56
 Operator : SY/MD
 Sample : P3692-23
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K94

Quant Time: Aug 28 06:14:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 28 05:52:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	664554	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	626242	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	306492	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	181904	29.923	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	59.840%	#	
7) Chloroethane-d5	1.529	69	166612	35.462	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	70.920%		
11) 1,1-Dichloroethene-d2	2.057	65	90840	35.553	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	71.100%		
21) 2-Butanone-d5	3.796	46	255370	108.478	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	108.480%		
24) Chloroform-d	4.243	84	406756	42.752	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	85.500%		
26) 1,2-Dichloroethane-d4	4.937	65	267809	44.698	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.400%		
32) Benzene-d6	4.954	84	812115	42.935	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	85.880%		
36) 1,2-Dichloropropane-d6	5.986	67	264397	44.628	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.260%		
41) Toluene-d8	7.240	98	717320	42.020	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	84.040%		
43) trans-1,3-Dichloroprop...	7.551	79	105247	38.399	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	76.800%		
47) 2-Hexanone-d5	8.024	63	165629	102.437	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	102.440%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	339686	47.276	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	94.560%		
66) 1,2-Dichlorobenzene-d4	11.555	152	279205	45.485	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	90.960%		
Target Compounds						Qvalue
19) 1,1-Dichloroethane	3.105	63	18284	2.046	ug/L	98
20) cis-1,2-Dichloroethene	3.806	96	44144	8.876	ug/L	94
34) Trichloroethene	5.838	95	17626	3.695	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082724\
Data File : VV037103.D
Acq On : 27 Aug 2024 21:56
Operator : SY/MD
Sample : P3692-23
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K94

Quant Time: Aug 28 06:14:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 28 05:52:22 2024
Response via : Initial Calibration

