

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082724\
 Data File : VW037104.D
 Acq On : 27 Aug 2024 22:18
 Operator : SY/MD
 Sample : P3693-06
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K40

Quant Time: Aug 28 06:14:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 28 05:52:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	656455	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	623611	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	298941	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	176108	29.327	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	58.660%#		
7) Chloroethane-d5	1.529	69	165511	35.663	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	71.320%		
11) 1,1-Dichloroethene-d2	2.056	65	88467	35.052	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	70.100%		
21) 2-Butanone-d5	3.796	46	260119	111.859	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	111.860%		
24) Chloroform-d	4.243	84	410798	43.710	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.420%		
26) 1,2-Dichloroethane-d4	4.937	65	270505	45.704	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.400%		
32) Benzene-d6	4.953	84	799636	42.454	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.900%		
36) 1,2-Dichloropropane-d6	5.986	67	263381	44.645	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.280%		
41) Toluene-d8	7.240	98	715741	42.104	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	84.200%		
43) trans-1,3-Dichloroprop...	7.551	79	105266	38.568	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	77.140%		
47) 2-Hexanone-d5	8.024	63	168406	104.594	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	104.590%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	346675	48.452	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	96.900%		
66) 1,2-Dichlorobenzene-d4	11.555	152	275739	46.055	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	92.100%		
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	3.809	96	19748	4.020	ug/L	94
33) Benzene	5.002	78	1089318	60.528	ug/L	100
34) Trichloroethene	5.838	95	13314	2.803	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082724\
Data File : VV037104.D
Acq On : 27 Aug 2024 22:18
Operator : SY/MD
Sample : P3693-06
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K40

Quant Time: Aug 28 06:14:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 28 05:52:22 2024
Response via : Initial Calibration









#20
 cis-1,2-Dichloroethene
 Concen: 4.020 ug/L
 RT: 3.809 min Scan# 80
 Delta R.T. 0.003 min
 Lab File: VV037104.D
 Acq: 27 Aug 2024 22:18




Instrument :
 MSVOA_V
 ClientSampleId :
 E1K40


Tgt Ion: 96 Resp: 19748
 Ion Ratio Lower Upper
 96 100
 61 123.2 89.9 166.9
 98 56.6 44.8 83.2



#33
 Benzene
 Concen: 60.528 ug/L
 RT: 5.002 min Scan# 1232
 Delta R.T. 0.000 min
 Lab File: VV037104.D
 Acq: 27 Aug 2024 22:18

Tgt Ion: 78 Resp: 1089318





#34
 Trichloroethene
 Concen: 2.803 ug/L
 RT: 5.838 min Scan# 14
 Delta R.T. 0.010 min
 Lab File: VV037104.D
 Acq: 27 Aug 2024 22:18

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K40

Tgt Ion: 95 Resp: 13314

Ion	Ratio	Lower	Upper
95	100		
97	65.3	44.4	82.5
132	91.5	64.9	120.5
130	100.2	69.4	129.0

