

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
 Data File : VV037124.D
 Acq On : 28 Aug 2024 10:59
 Operator : SY/MD
 Sample : P3693-22
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Aug 29 04:45:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 29 04:43:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	686092	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	657538	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	309240	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	138078	22.001	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	44.000%#
7) Chloroethane-d5	1.532	69	131336	27.077	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	54.160%#
11) 1,1-Dichloroethene-d2	2.056	65	70323	26.659	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.320%#
21) 2-Butanone-d5	3.796	46	187875	77.302	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.300%
24) Chloroform-d	4.243	84	332934	33.895	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.780%#
26) 1,2-Dichloroethane-d4	4.937	65	215889	34.901	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	69.800%#
32) Benzene-d6	4.953	84	638016	32.125	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	64.260%#
36) 1,2-Dichloropropane-d6	5.986	67	217070	34.896	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	69.800%#
41) Toluene-d8	7.239	98	556878	31.069	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	62.140%#
43) trans-1,3-Dichloroprop...	7.551	79	86288	29.983	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.960%#
47) 2-Hexanone-d5	8.024	63	115573	68.077	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	68.080%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	274540	36.391	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	72.780%
66) 1,2-Dichlorobenzene-d4	11.554	152	218354	35.256	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	70.520%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
Data File : VV037124.D
Acq On : 28 Aug 2024 10:59
Operator : SY/MD
Sample : P3693-22
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 29 04:45:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 29 04:43:38 2024
Response via : Initial Calibration

