

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
 Data File : VV037130.D
 Acq On : 28 Aug 2024 15:07
 Operator : SY/MD
 Sample : P3692-17
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K66

Quant Time: Aug 29 04:46:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 29 04:43:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	596279	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	577214	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	278292	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	268595	49.243	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 98.480%			
7) Chloroethane-d5	1.532	69	212268	50.353	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 100.700%			
11) 1,1-Dichloroethene-d2	2.057	65	114683	50.024	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 100.040%			
21) 2-Butanone-d5	3.796	46	237575	112.474	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 112.470%			
24) Chloroform-d	4.243	84	429260	50.283	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 100.560%			
26) 1,2-Dichloroethane-d4	4.937	65	272831	50.750	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 101.500%			
32) Benzene-d6	4.953	84	873017	50.075	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 100.160%			
36) 1,2-Dichloropropane-d6	5.986	67	279462	51.178	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 102.360%			
41) Toluene-d8	7.240	98	769664	48.916	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 97.840%			
43) trans-1,3-Dichloroprop...	7.551	79	114905	45.483	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 90.960%			
47) 2-Hexanone-d5	8.024	63	159968	107.340	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 107.340%			
56) 1,1,2,2-Tetrachloroeth...	10.146	84	341014	51.492	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 102.980%			
66) 1,2-Dichlorobenzene-d4	11.555	152	281262	50.463	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 100.920%			
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	3.812	96	15293	3.427	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
Data File : VV037130.D
Acq On : 28 Aug 2024 15:07
Operator : SY/MD
Sample : P3692-17
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K66

Quant Time: Aug 29 04:46:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 29 04:43:38 2024
Response via : Initial Calibration

