

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082824\
 Data File : VW037136.D
 Acq On : 28 Aug 2024 17:19
 Operator : SY/MD
 Sample : P3692-21DL 20X
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K81DL

Quant Time: Aug 29 04:47:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 29 04:43:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	599747	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	575273	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	269989	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	265071	48.316	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	96.640%		
7) Chloroethane-d5	1.529	69	208602	49.197	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	98.400%		
11) 1,1-Dichloroethene-d2	2.056	65	110795	48.049	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	96.100%		
21) 2-Butanone-d5	3.796	46	240324	113.118	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	113.120%		
24) Chloroform-d	4.246	84	425591	49.565	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.140%		
26) 1,2-Dichloroethane-d4	4.937	65	269542	49.848	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.700%		
32) Benzene-d6	4.953	84	862850	49.659	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.320%		
36) 1,2-Dichloropropane-d6	5.986	67	272473	50.066	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	100.140%		
41) Toluene-d8	7.239	98	750460	47.856	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	95.720%		
43) trans-1,3-Dichloroprop...	7.551	79	114085	45.311	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	90.620%		
47) 2-Hexanone-d5	8.024	63	165967	111.741	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	111.740%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	335756	50.869	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	101.740%		
66) 1,2-Dichlorobenzene-d4	11.554	152	270337	49.994	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	99.980%		
Target Compounds						Qvalue
5) Vinyl chloride	1.282	62	13017	2.433	ug/L	# 49
20) cis-1,2-Dichloroethene	3.809	96	264219	58.866	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
Data File : VV037136.D
Acq On : 28 Aug 2024 17:19
Operator : SY/MD
Sample : P3692-21DL 20X
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K81DL

Quant Time: Aug 29 04:47:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 29 04:43:38 2024
Response via : Initial Calibration

