

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
 Data File : VV037137.D
 Acq On : 28 Aug 2024 17:42
 Operator : SY/MD
 Sample : P3692-16DL 200X
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K65DL

Quant Time: Aug 29 04:48:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 29 04:43:38 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	561102	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	541563	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	249833	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	256976	50.066	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	100.140%	
7) Chloroethane-d5	1.532	69	206985	52.178	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	104.360%	
11) 1,1-Dichloroethene-d2	2.056	65	109909	50.947	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	101.900%	
21) 2-Butanone-d5	3.796	46	245611	123.569	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	123.570%	
24) Chloroform-d	4.243	84	415474	51.720	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	103.440%	
26) 1,2-Dichloroethane-d4	4.937	65	269907	53.353	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	106.700%	
32) Benzene-d6	4.953	84	849534	51.936	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	103.880%	
36) 1,2-Dichloropropane-d6	5.985	67	268405	52.389	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	104.780%	
41) Toluene-d8	7.239	98	743414	50.358	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	100.720%	
43) trans-1,3-Dichloroprop...	7.551	79	111680	47.117	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	94.240%	
47) 2-Hexanone-d5	8.024	63	160875	115.054	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	115.050%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	337450	54.308	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	108.620%	
66) 1,2-Dichlorobenzene-d4	11.558	152	272140	54.388	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	108.780%	
Target Compounds						Qvalue
5) Vinyl chloride	1.281	62	34076	6.807	ug/L	84
20) cis-1,2-Dichloroethene	3.805	96	498422	118.693	ug/L	100
34) Trichloroethene	5.825	95	301315	73.043	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
Data File : VV037137.D
Acq On : 28 Aug 2024 17:42
Operator : SY/MD
Sample : P3692-16DL 200X
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K65DL

Quant Time: Aug 29 04:48:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 29 04:43:38 2024
Response via : Initial Calibration

