

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW082824\
 Data File : VW037131.D
 Acq On : 28 Aug 2024 15:29
 Operator : SY/MD
 Sample : P3692-18
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1K69

Quant Time: Aug 29 04:46:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 29 04:43:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	594453	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	568763	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	264577	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	273533	50.302	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.600%		
7) Chloroethane-d5	1.532	69	212598	50.586	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 101.180%		
11) 1,1-Dichloroethene-d2	2.060	65	116351	50.908	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.820%		
21) 2-Butanone-d5	3.799	46	247871	117.709	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 117.710%		
24) Chloroform-d	4.246	84	434010	50.996	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.000%		
26) 1,2-Dichloroethane-d4	4.937	65	273808	51.088	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.180%		
32) Benzene-d6	4.953	84	884468	51.486	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.980%		
36) 1,2-Dichloropropane-d6	5.985	67	279284	51.905	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 103.820%		
41) Toluene-d8	7.239	98	779480	50.276	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 100.560%		
43) trans-1,3-Dichloroprop...	7.551	79	114548	46.016	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 92.040%		
47) 2-Hexanone-d5	8.024	63	162934	110.954	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 110.950%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	343623	52.657	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 105.320%		
66) 1,2-Dichlorobenzene-d4	11.554	152	282620	53.335	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.680%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082824\
Data File : VV037131.D
Acq On : 28 Aug 2024 15:29
Operator : SY/MD
Sample : P3692-18
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1K69

Quant Time: Aug 29 04:46:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 29 04:43:38 2024
Response via : Initial Calibration

