

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082924\
 Data File : VV037163.D
 Acq On : 29 Aug 2024 17:17
 Operator : SY/MD
 Sample : VIBLK370
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK370

Quant Time: Aug 29 23:44:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 29 23:38:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	607210	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	586735	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	272087	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	238092	42.865	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.720%
7) Chloroethane-d5	1.532	69	195919	45.638	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.280%
11) 1,1-Dichloroethene-d2	2.056	65	105078	45.009	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.020%
21) 2-Butanone-d5	3.796	46	208676	97.014	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.010%
24) Chloroform-d	4.243	84	407812	46.911	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.820%
26) 1,2-Dichloroethane-d4	4.937	65	262783	48.001	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.000%
32) Benzene-d6	4.953	84	829642	46.815	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.640%
36) 1,2-Dichloropropane-d6	5.985	67	267635	48.217	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.440%
41) Toluene-d8	7.239	98	731545	45.739	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.480%
43) trans-1,3-Dichloroprop...	7.551	79	109153	42.505	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.020%
47) 2-Hexanone-d5	8.024	63	138149	91.195	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.190%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	319153	47.409	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.820%
66) 1,2-Dichlorobenzene-d4	11.554	152	267015	48.999	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082924\
Data File : VV037163.D
Acq On : 29 Aug 2024 17:17
Operator : SY/MD
Sample : VIBLK370
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK370

Quant Time: Aug 29 23:44:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 29 23:38:00 2024
Response via : Initial Calibration

