

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090624\
 Data File : VV037214.D
 Acq On : 06 Sep 2024 10:59
 Operator : SY/MD
 Sample : P3842-06
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Sep 06 23:22:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 06 23:21:35 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	593102	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	576204	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	263386	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	225415	41.548	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	83.100%	
7) Chloroethane-d5	1.532	69	190752	45.492	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	90.980%	
11) 1,1-Dichloroethene-d2	2.056	65	101919	44.695	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	89.380%	
21) 2-Butanone-d5	3.789	46	178621	85.017	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	85.020%	
24) Chloroform-d	4.240	84	419677	49.424	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.840%	
26) 1,2-Dichloroethane-d4	4.934	65	264401	49.445	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.880%	
32) Benzene-d6	4.950	84	847066	48.672	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.340%	
36) 1,2-Dichloropropane-d6	5.979	67	277088	50.832	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.660%	
41) Toluene-d8	7.236	98	725674	46.201	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	92.400%	
43) trans-1,3-Dichloroprop...	7.548	79	114466	45.389	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.780%	
47) 2-Hexanone-d5	8.024	63	106889	71.849	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	71.850%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	292867	44.299	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	88.600%	
66) 1,2-Dichlorobenzene-d4	11.554	152	271522	51.472	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.940%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090624\
Data File : VV037214.D
Acq On : 06 Sep 2024 10:59
Operator : SY/MD
Sample : P3842-06
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Sep 06 23:22:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 06 23:21:35 2024
Response via : Initial Calibration

