

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090624\
 Data File : VV037211.D
 Acq On : 06 Sep 2024 09:17
 Operator : SY/MD
 Sample : VV0906MBL01
 Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK200

Quant Time: Sep 06 23:22:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 06 23:21:35 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	657138	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	619563	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	293327	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	282647	47.020	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.040%
7) Chloroethane-d5	1.532	69	220783	47.523	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.040%
11) 1,1-Dichloroethene-d2	2.060	65	120534	47.707	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.420%
21) 2-Butanone-d5	3.789	46	179326	77.035	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.040%
24) Chloroform-d	4.239	84	440707	46.843	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.680%
26) 1,2-Dichloroethane-d4	4.931	65	285771	48.234	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.460%
32) Benzene-d6	4.950	84	933665	49.893	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.780%
36) 1,2-Dichloropropane-d6	5.979	67	298357	50.904	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.800%
41) Toluene-d8	7.236	98	826589	48.943	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.880%
43) trans-1,3-Dichloroprop...	7.548	79	122652	45.231	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.460%
47) 2-Hexanone-d5	8.021	63	109382	68.379	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	68.380%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	295932	41.630	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.260%
66) 1,2-Dichlorobenzene-d4	11.554	152	293337	49.932	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090624\
Data File : VV037211.D
Acq On : 06 Sep 2024 09:17
Operator : SY/MD
Sample : VV0906MBL01
Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK200

Quant Time: Sep 06 23:22:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 06 23:21:35 2024
Response via : Initial Calibration

